

AGATA S001

60Co source 1173.23 1332.51(Calibration)

Vop +4000 V

Shaping Amp. Intertechnique 7243E

t\_shaping 4  $\mu$ s

Date 16.06.10

Meas. time 600s

Electrode:	FWHM (keV)	F10/F2
Core	2.63	1.862
A1	2.21	1.937
A2		
A3		
A4		
A5		
A6		
B1	2.18	1.884
B2	2.02	1.899
B3	2.08	1.858
B4	2.27	1.927
B5	2.29	1.875
B6	2.27	1.975
C1		
C2		
C3		
C4		
C5		
C6		
D1		
D2		
D3		
D4		
D5		
D6		
E1		
E2		
E3		
E4		
E5		
E6		
F1		
F2		
F3		
F4		
F5		
F6		