


TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6178



TEST RESULTS


Nuclide : Cs-137 (662 keV)
Energy resolution : 9.6 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann

Signature :



TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6179



TEST RESULTS

Nuclide : Cs-137 (662 keV)
Energy resolution : 9.7 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann

Signature :






Dedicated
Scintillation Detectors

TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6180



TEST RESULTS

Nuclide : Cs-137 (662 keV)
Energy resolution : 9.6 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann


Signature :



Dedicated
Scintillation Detectors

TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6181



TEST RESULTS

Nuclide : Cs-137 (662 keV)
Energy resolution : 9.6 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann


Signature :



Dedicated
Scintillation Detectors

TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6182



TEST RESULTS

Nuclide : Cs-137 (662 keV)
Energy resolution : 9.8 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann


Signature :



Dedicated
Scintillation Detectors

TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6183



TEST RESULTS

Nuclide : Cs-137 (662 keV)
Energy resolution : 10.0 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann

Signature :






Dedicated
Scintillation Detectors

TEST SHEET

Detector type : VS-1368-11
Crystal : BGO
Read-out : Hamamatsu 3 1/2" Type R10233-100
Serial number : S2AA6184



TEST RESULTS

Nuclide : Cs-137 (662 keV)
Energy resolution : 10.0 %
Noise : N/A
Peak to valley : N/A
GEE : N/A
High voltage : + 980 V
Date : 16 March 2018
Remarks : N/A
Tested by : Eric Schlamann

Signature :